



Valid from 30 October 2019  
to 29 October 2022  
Issued on 30 October 2019

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



**ISO/IEC 17025  
TL 006-01**

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 006-01

**Lindel Industrial Laboratories Limited  
Pattiwila Road  
Sapugaskanda  
Makola.**

**Scope of Accreditation:** Performing Chemical testing on Water.

The laboratory is accredited for the tests in page 02.

SI	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Waste Water/ Potable Water	COD	APHA 5220D – 23 <sup>rd</sup> Version (Vial Method)	5 – 150 ppm 151 – 1500 ppm
02	Waste Water/ Potable Water	Conductivity	APHA 2510B – 23 <sup>rd</sup> Version	10-150 $\mu$ S/cm 151-1500 $\mu$ S/cm 1.5 – 1288 mS/cm
03	Waste Water/ Potable Water	pH	APHA 4500H <sup>+</sup> B- – 23 <sup>rd</sup> Version	3.0 – 11.0
04	Waste Water/ Potable Water	TDS	APHA 2540C – 23 <sup>rd</sup> Version	50 – 250 ppm 251 – 4000 ppm
05	Waste Water/ Potable Water	TS	APHA 2540B – 23 <sup>rd</sup> Version	15 - 50 ppm 51 - 250 ppm
06	Waste Water/ Potable Water	TSS	APHA 2540D – 23 <sup>rd</sup> Version	15 – 100 ppm 101 – 500 ppm
07	Waste Water	Oil & Grease	APHA 5520B – 23 <sup>rd</sup> Version	10 – 50 ppm 4 – 9 ppm
08	Waste Water	BOD	APHA 5210B – 23 <sup>rd</sup> Version	10 – 30 ppm 31 – 500 ppm
09	Potable Water	Alkalinity	APHA 2320B – 23 <sup>rd</sup> Version	2.0 – 200.0
10	Potable Water	Hardness	APHA 2340C – 23 <sup>rd</sup> Version	5 – 50 ppm
11	Potable Water	Nitrate	APHA 4500 NO <sub>3</sub> <sup>-</sup> B – 23 <sup>rd</sup> Version	0.5 – 5.0 ppm

Director/CEO  
Sri Lanka Accreditation Board for Conformity Assessment