



Valid from 15 February 2017
to 16 November 2017
Issued on 05 June 2017

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
TL 052-01

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 052-01

Regional Laboratory-Western Production (Ambatale)
National Water Supply & Drainage Board
Udumulla Road
Mulleriyawa New Town

Scope of Accreditation: Performing chemical testing on pH, Conductivity, Turbidity and Sampling of water as per APHA methods.

The laboratory is accredited for the following tests.

Sl No.	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Water	Sampling of Water	APHA 1060: 21 st Edition	N/A
		pH	APHA 4500-H ⁺ B: 21 st Edition	2-12
		Conductivity	APHA 2510 B: 21 st Edition	0-400 (μ s/cm)
		Turbidity	APHA 2130 B: 21 st Edition	0.10-10.00 (NTU)
2.0-25.0 (NTU)				

Deputy Director (Accreditation)
Sri Lanka Accreditation Board for Conformity Assessment