



Valid from 30 October 2019
to 30 November 2020
Issued on 15 November 2019

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
TL 052 - 05

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity

Assessment Accreditation Number : TL 052 – 05

Regional Laboratory- Jaffna
National Water Supply and Drainage Board
No.46, Sivan Pannai Road
Jaffna

Scope of Accreditation: Performing Chemical Testing on Water as per the
Test Methods appearing in this Schedule.

The laboratory is accredited for the following tests.

SI	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
1	Water	pH at 25±2 ⁰ C	APHA 4500 H ⁺ B 23 rd Edition	0.5 – 12
		Conductivity	APHA 2510 B 23 rd Edition	0.2 – 150,000 µs/cm
		Turbidity	APHA 2130 B 23 rd Edition	0.05-4000 NTU
		Chloride	APHA 4500 Cl ⁻ B 23 rd Edition	0-12000 mg/L
		Total Hardness as CaCO ₃	APHA 2340 C 23 rd Edition	0-8000 mg/L

Director /CEO
Sri Lanka Accreditation Board for Conformity Assessment