



Valid from 30 October 2019
to 30 November 2020
Issued on 15 November 2019

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
TL 052 - 07

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity

Assessment Accreditation Number : TL 052 – 07

Regional Laboratory- Vavuniya
National Water Supply and Drainage Board
A9 Road, Vavuniya

Scope of Accreditation: Performing Chemical Testing on Water as per the
Test Methods appearing in this Schedule.

The laboratory is accredited for the following tests.

| SI | Product(s) / Material of test | Specific tests performed | Test Method / Standard against which tests are performed | Limits of detection |
|----|-------------------------------------|--|---|---------------------|
| 1 | Water | pH at 25±2 ⁰ C | APHA 4500 H ⁺ B 23 rd Edition | 0.22 |
| | | Conductivity | APHA 2510 B 23 rd Edition | 8.9 µs/cm |
| | | Turbidity | APHA 2130 B 23 rd Edition | 0.3 NTU |
| | | Chloride | APHA 4500 Cl ⁻ B 23 rd Edition | 1.0 mg/L |
| | | Total Hardness as CaCO ₃ | APHA 2340 C 23 rd Edition | 2.0 mg/L |

Director /CEO
Sri Lanka Accreditation Board for Conformity Assessment